

SPECIFICATION SHEET

Product Name

Buffer Solution pH 6.00 ± 0.01 at 20°C

(Potassium di-hydrogen Orthophosphate, Disodium Hydrogen Orthophosphate)

Product Number

: 908.B06

Brand

ISOLAB

CAS Number

-

Formula

_

Formula Weight

-

TEST NAME	SPECIFICATION	UNIT
pH	5.99 - 6.01 at 20°C	-
Measured with a combination glass electrode after multiple point calibration with reference		

Measured with a combination glass electrode after multiple point calibration with reference materials. It is certified traceable to the following National Institute of Standards and Technology (USA), SRM 185i Potassium Hydrogen Phthalate, SRM 186-I-g Potassium Dihydrogen Phosphate and SRM 186-II-g Disodium Hydrogen Phosphate.

